## Certificate of Compliance

No. 0B170605.SUSUS25

Test Report no. ENC170523GZ40L1, ENC170523GZ40E1

Certificate's

Holder:

SUS Advancing Technology Co., Ltd.

3/F, #6 Jiahe Wanggang Luduan, West Jichang Road,

Baiyun District, Guangzhou 510440

**Certification ECM** 

Mark:



Product:

Photo Dynamic Therapy System

Brand:

**SUSLASER** 

Model(s):

BL001, BL002, BL003, BL004, BL005, BL006, BL007, BL008, BL009, BL010, BL020, BL030, BL040, BL050

Verification to:

Standard:

EN 60335-1:2012+A11:2014+AC:2014+

A12:2017, EN 60335-2-27:2013, EN 62233:2008,

EN 55014-1:2017, EN 55014-2:2015, EN 61000-3-2:2014, EN 61000-3-3:2013

related to CE Directive(s): 2014/35/EU (Low Voltage)

2014/30/EU (Electromagnetic Compatibility)

Remark: The product(s) has been verified on a voluntary basis. The product(s) satisfies the requirements of the Certification Mark of ECM, in reference to the above listed Standard(s). The above Compliance Mark can be affixed on the product(s) accordingly to the ECM regulation about its release and its use. The regulation can be found at www.entecerma.it. This Certificate of Compliance can be checked for validity at www.entecerma.it

This verification doesn't imply assessment of the production of the product(s).

Additional information, clarification about the CE Marking:



We attest that a TCF for the CE Marking process is in place. Whereas the Manufacturer is Responsible to start the € Marking Certification Procedure and to perform all the necessary activities, as required by the Directive before placing the **C**€ Mark on the product(s).

Date of issue 05 June 2017

Expiry date 04 June 2022



**Deputy Manager** Amanda Payne